



ANALYSIS REPORT

Report No. : **2226759E-R1** Report Date : 30/09/2022
Applicant : BRBEN TEKST L SAN. VE T C. A. .
Address : 2. OSB 83207 Nolu Cad. No: 2/10 ehitkamil/Gaziantep/Turkey
Sample : Overalls BRTLM-001 CAT.III Type 5-B/6-B Protective Suit Tuta Protettiva
CAT.III Type 5-B/6-B -BR MED
Sample Package : Original poly packing
Sample Amount : 2 pieces
Sampling Point : -
Sampling Method : -
Sampling Date : -
Sample Lot No. : BRTLM2021/0036
Production Date : 20/09/2021
Packing Date : -
Expire Date : 20/09/2024
Producer Company : Brben Tekstil San. ve Tic. A. .
Product No : -
Supplier Number : -
Sample Receiving Time : 21/09/2022 16:15:00
Analysis Beginning Time : 21/09/2022 16:30:00
Analysis Completion Time : 29/09/2022

Following analysis results were obtained from the specimen which was delivered by cargo to Çevre Laboratory

Parameters	Unit	Finding	Method	Information
Microbial Penetration - Wet Bacterium				
Test Spicemen 1 - Colony Count	cfu	3	ISO 22610	(*) 154, 180
Test Spicemen 2 - Colony Count	cfu	4	ISO 22610	(*) 154, 180
Test Spicemen 3 - Colony Count	cfu	4	ISO 22610	(*) 154, 180
Test Spicemen 4 - Colony Count	cfu	3	ISO 22610	(*) 154, 180
Test Spicemen 5 - Colony Count	cfu	4	ISO 22610	(*) 154, 180
Test Spicemen 1 - Percentage of Penetration	%	0,06	ISO 22610	(*) 154
Test Spicemen 2 - Percentage of Penetration	%	0,08	ISO 22610	(*) 154
Test Spicemen 3 - Percentage of Penetration	%	0,08	ISO 22610	(*) 154
Test Spicemen 4 - Percentage of Penetration	%	0,06	ISO 22610	(*) 154

e-signed
Sevinç ÖCAL
Assistant Laboratory Responsible of
Microbiology Laboratory

e-signed
Approved by
30/09/2022
Ömer Yasin BALIK
Laboratory Manager

ANALYSIS REPORT

Report No. : 2226759E-R1

Report Date : 30/09/2022

AB-0363-T

2226759E-R1

09-22

Following analysis results were obtained from the specimen which was delivered by cargo to Çevre Laboratory

Parameters	Unit	Finding	Method	Information
Test Spicemen 5 - Percentage of Penetration	%	0,08	ISO 22610	(*) 154
Average Penetration Percentage	%	0,07	ISO 22610	(*)
Bacillus atrophaeus Concentration	spores/mL	5,3*10 ³	ISO 22610	(*)

Method ISO : International Organization for Standardization

Information 154 : Test Conditions : 65±5 relative humidity and 20±2°C minimum 24 hours
The distance to the distance agar-to-brim is 3.0 mm.
25 cm x 25 cm 5 test pieces were used.
The tests were carried out from the outside of the sample.
ATCC 9372 Bacillus atrophaeus spore suspension was used.
Incubator Control <4 cfu
Test Environment Control <25 cfu

180 : EN 14126 standard provides Class 1 values according to Table 2.

R1 : This code means correction at this report 30/09/2022 dated and 2226759E numbered report is invalid. Based on the customer's request 'Report Format' has been changed and the report has been revised. Our report has been rearranged as new report number is 2226759E-R1.

Note

1. Çevre End. Analiz Lab. Hiz. Tic. A. . accredited by TÜRKAK under registration number [AB-0363-T] for [TS EN ISO/IEC 17025] as test laboratory". Turkish Accreditation Agency (TURKAK) is a signatory to the European co-operation for Accreditation (EA) Multilateral Agreement (MLA) and to the International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Arrangement (MRA) for the recognition of test reports.
2. When a conformity assessment is requested, legal regulations, standards or the decision rule agreed with the customer are applied by us after a risk assessment and the application method is specified in the information section of the report. The Simple Decision Rule applies where conformity assessment is requested without taking into account the measurement uncertainty.
3. Uncertainties stated in the report are expanded uncertainty (k=2, 95%). Total Uncertainty of Measurement includes the uncertainty from sampling. It is valid when the sample is taken by us.
4. The definitional information included in the analysis report and affecting the validity of the results has been declared by the customer. Our laboratory is not responsible for any losses/legal obligations that may occur due to the accuracy and use of this information.
5. Analysis report covers samples/sampling that comes to the laboratory.
6. This report and results don't not be copied and printed partially or completely without permission of Çevre Industrial Analysis Laboratory for any commercial and advertising purposes.
7. This report shall not be used for legal/administrative procedures and advertising purposes.
8. The test report without sign is not valid.
9. (*) This parameter is covered by our accreditation scope.

End of Report

e-signed

Sevinç ÖCAL

Assistant Laboratory Responsible of
Microbiology Laboratory

e-signed

Approved by

30/09/2022

Ömer Yasin BALIK
Laboratory Manager