



ANALYSIS REPORT

Report No. : **2323713E** Report Date : 30/10/2023
Applicant : BRBEN TEKSTİL SAN. VE TİC. A.Ş.
Address : 2. OSB 83207 Nolu Cad. No: 2/10 Ehitkamil/Gaziantep/Turkey
Sample : Disposable Coverall (L) - BR Med BRTLM-001 / CAT.III Type 5-B/6-B Protective Suit
Sample Package : Original poly packing
Sample Amount : 2 pieces
Sampling Point : -
Sampling Method : -
Sampling Date : -
Sample Lot No. : BR2023/5464
Sample Carrying Conditions / Preservation Technique : Environment temperature condition
Production Date : 01/09/2023
Packing Date : -
Expire Date : 01/09/2026
Producer Company : Brben Tekstil San. ve Tic. A.Ş.
Sample Receiving Time : 19/10/2023 13:30:00
Analysis Beginning Time : 19/10/2023 13:45:00
Analysis Completion Time : 27/10/2023

Following analysis results were obtained from the specimen which was delivered by cargo to Çevre Laboratory;

Parameters	Unit	Finding	Method	Information
Microbial Penetration - Dry Bacterium				
Microbial Penetration - Dry Bacterium	log cfu	<1	ISO 22612	150, 151
Test Spicemen 1 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 2 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 3 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 4 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 5 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 6 - Colony Count	cfu	<1	ISO 22612	

Fatma M EK
Microbiology Laboratory Responsible

Approved by
20/11/2023
Ömer Yasin BALIK
Laboratory Manager

ANALYSIS REPORT

Report No. : 2323713E

Report Date : 30/10/2023

Following analysis results were obtained from the specimen which was delivered by cargo to Çevre Laboratory;

Parameters	Unit	Finding	Method	Information
Test Spicemen 7 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 8 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 9 - Colony Count	cfu	<1	ISO 22612	
Test Spicemen 10 - Colony Count	cfu	<1	ISO 22612	
Ortalama Koloni Sayısı	cfu	<1	ISO 22612	
Negative Control Count 1	cfu	<1	ISO 22612	
Negative Control Count 2	cfu	<1	ISO 22612	
Talc Concentration	cfu/g	3,1*10 ⁸	ISO 22612	

MU: Measurement Uncertainty

Method ISO : International Organization for Standardization

Information 150 : Test Conditions : 65±5 relative humidity and 20±2°C
ATCC 9372 Bacillus subtilis spores were used in the concentration of ethyl alcohol.
200 mm x 200 mm 12 test pieces used
The vibrator was operated in an air flow with a vibration frequency of 20800 per minute.

151 : EN 14126 standard provides Class 3 values according to Table 4.

Note

1. When a conformity assessment is requested, legal regulations, standards or the decision rule agreed with the customer are applied by us after a risk assessment and the application method is specified in the information section of the report. The Simple Decision Rule applies where conformity assessment is requested without taking into account the measurement uncertainty.
2. The definitional information included in the analysis report and affecting the validity of the results has been declared by the customer. Our laboratory is not responsible for any losses/legal obligations that may occur due to the accuracy and use of this information.
3. Analysis report covers samples/sampling that comes to the laboratory.
4. This report and results don't not be copied and printed partially or completely without permission of Çevre Industrial Analysis Laboratory for any commercial and advertising purposes.
5. This report shall not be used for legal/administrative procedures and advertising purposes.
6. The test report without sign is not valid.

End of Report



Fatma M EK
Microbiology Laboratory Responsible



Approved by
20/11/2023
Ömer Yasin BALIK
Laboratory Manager